


GHRP-6 HPLC

Test time	:05/06/2018, 10:12	Product Name	:GHRP-6
Analyst	:LZH	Batch number	:US20180506
Column	:GSA46250BI017013, C18, 4.6*250mm, 5um	Flow Rate	:1.0ml/min
Solvent A	:100% Acetonitrile	Wavelength	:220nm
Solvent B	:0.1% Trifluoroacetic in 100% Water	Temperature	:37°C
Gradient	:	Concentration	:1mg/ml
	A B	Sample Volume	:10ul
	0.0min 5% 95%		
	25.0min 95% 5%		
	25.1min 100% 0%		
	30.0min Stop		



Number	Stay time	Concentration	Peak area	Peak height
1	13.425	0.1174	8567	2563
2	13.613	99.07	7230454	1356731
3	14.076	0.1903	13890	2666
4	14.544	0.2053	14982	3608
5	14.743	0.4116	30043	5598
Total		100	7297936	1371166