


## IGF-1 LR3 HPLC

Test time	:05/07/2018, 13:29	Product Name	:IGF-1 LR3
Analyst	:LZH	Batch number	:US20180507
Column	:GSA46250BI017013, C18, 4.6*250mm, 5um	Flow Rate	:1.0ml/min
Solvent A	:100% Acetonitrile	Wavelength	:220nm
Solvent B	:0.1% Trifluoroacetic in 100% Water	Temperature	:37°C
Gradient	:	Concentration	:1mg/ml
	A B	Sample Volume	:20ul
	0.0min 30% 70%		
	25.0min 60% 40%		
	25.1min 100% 0%		
	30.0min Stop		



Number	Stay time	Concentration	Peak area	Peak height
1	9.806	0.8745	32627	4949
2	10.949	0.6445	24045	3127
3	11.413	98.48	3674197	306030
Total		100	3730869	314106